

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081319\
 Data File : VX011548.D
 Acq On : 13 Aug 2019 19:58
 Operator : JC/SP
 Sample : K4266-12
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 R1-R2-C1-(0)

Quant Time: Aug 14 02:39:19 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
 Quant Title : SW846 8260
 QLast Update : Fri Aug 02 01:55:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	192003	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	292288	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	269269	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	118629	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	108965	56.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.70%	
35) Dibromofluoromethane	5.49	113	89730	47.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.04%	
50) Toluene-d8	8.71	98	348338	48.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.72%	
62) 4-Bromofluorobenzene	11.13	95	117701	45.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.54%	
Target Compounds						
16) Acetone	2.43	43	7658	8.321	ug/l	99
20) Methylene Chloride	2.85	84	7338	3.641	ug/l	98
43) Isopropyl Acetate	6.44	43	32858	8.029	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081319\
Data File : VX011548.D
Acq On : 13 Aug 2019 19:58
Operator : JC/SP
Sample : K4266-12
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
R1-R2-C1-(0)

Quant Time: Aug 14 02:39:19 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X080119W.M
Quant Title : SW846 8260
QLast Update : Fri Aug 02 01:55:00 2019
Response via : Initial Calibration

