

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081320\
 Data File : VX017872.D
 Acq On : 13 Aug 2020 13:55
 Operator : JC/SP
 Sample : L3643-07
 Misc : 25.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BG0A9

Quant Time: Aug 14 07:58:17 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\SOMXTR080720WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Aug 14 07:35:30 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.83	114	261087	5.00	ug/L	0.00
28) Chlorobenzene-d5	10.10	117	230164	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	12.06	152	104845	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	99890	5.54	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	110.80%
7) Chloroethane-d5	1.69	69	77649	5.42	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	108.40%
11) 1,1-Dichloroethene-d2	2.34	63	138124	3.72	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	74.40%
20) 2-Butanone-d5	4.53	46	228530	47.70	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	95.40%
24) Chloroform-d	5.14	84	179321	4.90	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.00%
26) 1,2-Dichloroethane-d4	6.03	65	100915	5.02	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.40%
32) Benzene-d6	6.05	84	341329	5.40	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	108.00%
36) 1,2-Dichloropropane-d6	7.37	67	102877	5.27	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	105.40%
41) Toluene-d8	8.70	98	283047	4.70	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.00%
45) trans-1,3-Dichloropropene-	8.99	79	34183	4.04	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	80.80%
48) 2-Hexanone-d5	9.43	63	154989	44.09	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	88.18%
59) 1,1,2,2-Tetrachloroethane-	11.23	84	69397	4.67	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.40%
65) 1,2-Dichlorobenzene-d4	12.36	152	106667	5.28	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	105.60%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081320\
Data File : VX017872.D
Acq On : 13 Aug 2020 13:55
Operator : JC/SP
Sample : L3643-07
Misc : 25.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BG0A9

Quant Time: Aug 14 07:58:17 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\SOMXTR080720WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Aug 14 07:35:30 2020
Response via : Initial Calibration

