

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081321\
 Data File : VX023679.D
 Acq On : 13 Aug 2021 10:32
 Operator : JC/MD
 Sample : VX0813MBL01
 Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0813MBL01

Quant Time: Aug 14 05:21:41 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	145453	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	247545	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	221457	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	82212	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	101227	55.791	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	111.580%
35) Dibromofluoromethane	5.391	113	81339	52.553	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.100%
50) Toluene-d8	8.653	98	299852	52.730	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	105.460%
62) 4-Bromofluorobenzene	11.086	95	95336	49.304	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.600%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081321\
Data File : VX023679.D
Acq On : 13 Aug 2021 10:32
Operator : JC/MD
Sample : VX0813MBL01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0813MBL01

Quant Time: Aug 14 05:21:41 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

