

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081321\
 Data File : VX023686.D
 Acq On : 13 Aug 2021 13:28
 Operator : JC/MD
 Sample : PB138358ZHE#03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB138358ZHE#03

Quant Time: Aug 14 05:24:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	145806	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	249571	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	235308	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	94942	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	101837	55.991	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	111.980%
35) Dibromofluoromethane	5.397	113	80078	51.318	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.640%
50) Toluene-d8	8.653	98	311194	54.281	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	108.560%
62) 4-Bromofluorobenzene	11.085	95	104969	53.845	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	107.680%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	27836	31.199	ug/l #	87
20) Methylene Chloride	2.794	84	11382	6.558	ug/l #	85
43) Isopropyl Acetate	6.355	43	23226	5.266	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081321\
Data File : VX023686.D
Acq On : 13 Aug 2021 13:28
Operator : JC/MD
Sample : PB138358ZHE#03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB138358ZHE#03

Quant Time: Aug 14 05:24:34 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

