

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081321\
 Data File : VX023687.D
 Acq On : 13 Aug 2021 13:52
 Operator : JC/MD
 Sample : PB138358ZHE#04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB138358ZHE#04

Quant Time: Aug 14 05:24:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	148332	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	253771	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	239596	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	97069	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	105185	56.847	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	113.700%
35) Dibromofluoromethane	5.397	113	82511	52.002	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.000%
50) Toluene-d8	8.653	98	315996	54.206	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	108.420%
62) 4-Bromofluorobenzene	11.085	95	107791	54.377	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	108.760%
Target Compounds						Qvalue
16) Acetone	2.386	43	27992	30.840	ug/l	90
20) Methylene Chloride	2.794	84	11650	6.599	ug/l #	86
43) Isopropyl Acetate	6.348	43	23110	5.153	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081321\
Data File : VX023687.D
Acq On : 13 Aug 2021 13:52
Operator : JC/MD
Sample : PB138358ZHE#04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB138358ZHE#04

Quant Time: Aug 14 05:24:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

