

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081321\
 Data File : VX023690.D
 Acq On : 13 Aug 2021 15:02
 Operator : JC/MD
 Sample : PB138358ZHE#07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB138358ZHE#07

Quant Time: Aug 14 05:25:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	154229	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	264244	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	254175	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	103588	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	107004	55.619	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 111.240%			
35) Dibromofluoromethane	5.397	113	85297	51.627	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 103.260%			
50) Toluene-d8	8.653	98	330496	54.446	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 108.900%			
62) 4-Bromofluorobenzene	11.085	95	115939	56.170	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 112.340%			
Target Compounds						Qvalue
16) Acetone	2.386	43	27041	28.653	ug/l	93
20) Methylene Chloride	2.794	84	11057	6.023	ug/l #	90
43) Isopropyl Acetate	6.348	43	23023	4.930	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081321\
Data File : VX023690.D
Acq On : 13 Aug 2021 15:02
Operator : JC/MD
Sample : PB138358ZHE#07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB138358ZHE#07

Quant Time: Aug 14 05:25:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

