

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081321\
 Data File : VX023691.D
 Acq On : 13 Aug 2021 15:26
 Operator : JC/MD
 Sample : PB138358ZHE#08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB138358ZHE#08

Quant Time: Aug 14 05:26:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	144103	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	246722	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	240905	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	95928	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	102725	57.147	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	114.300%
35) Dibromofluoromethane	5.397	113	79836	51.754	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.500%
50) Toluene-d8	8.653	98	312934	55.214	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	110.420%
62) 4-Bromofluorobenzene	11.085	95	106636	55.332	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	110.660%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	26590	30.155	ug/l	90
20) Methylene Chloride	2.794	84	10581	6.169	ug/l	91
43) Isopropyl Acetate	6.354	43	22011	5.048	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081321\
Data File : VX023691.D
Acq On : 13 Aug 2021 15:26
Operator : JC/MD
Sample : PB138358ZHE#08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB138358ZHE#08

Quant Time: Aug 14 05:26:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

