

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081321\
 Data File : VX023697.D
 Acq On : 13 Aug 2021 17:47
 Operator : JC/MD
 Sample : PB138358ZHE#14
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB138358ZHE#14

Quant Time: Aug 14 05:28:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	148274	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	252848	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	242063	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	96106	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	105366	56.967	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	113.940%
35) Dibromofluoromethane	5.397	113	82266	52.037	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.080%
50) Toluene-d8	8.653	98	318606	54.853	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	109.700%
62) 4-Bromofluorobenzene	11.085	95	107207	54.280	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	108.560%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	26758	29.492	ug/l	99
20) Methylene Chloride	2.794	84	11244	6.371	ug/l #	90
43) Isopropyl Acetate	6.354	43	22027	4.929	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081321\
Data File : VX023697.D
Acq On : 13 Aug 2021 17:47
Operator : JC/MD
Sample : PB138358ZHE#14
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB138358ZHE#14

Quant Time: Aug 14 05:28:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

