

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081321\
 Data File : VX023709.D
 Acq On : 13 Aug 2021 23:41
 Operator : JC/MD
 Sample : PB138358ZHE#20
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB138358ZHE#20

Quant Time: Aug 14 05:32:41 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	148981	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	259408	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	245940	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	95525	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	105867	56.967	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 113.940%	
35) Dibromofluoromethane	5.397	113	82981	51.162	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 102.320%	
50) Toluene-d8	8.653	98	322861	54.180	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 108.360%	
62) 4-Bromofluorobenzene	11.085	95	108145	53.370	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 106.740%	
Target Compounds						Qvalue
16) Acetone	2.386	43	28580	31.351	ug/l	90
20) Methylene Chloride	2.794	84	12587	7.098	ug/l	91
43) Isopropyl Acetate	6.349	43	23162	5.052	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081321\
Data File : VX023709.D
Acq On : 13 Aug 2021 23:41
Operator : JC/MD
Sample : PB138358ZHE#20
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB138358ZHE#20

Quant Time: Aug 14 05:32:41 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

