

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081321\
 Data File : VX023710.D
 Acq On : 14 Aug 2021 00:04
 Operator : JC/MD
 Sample : M3371-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 JET-GROUT

Quant Time: Aug 14 05:33:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	151028	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	261715	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	236688	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	89092	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	107719	57.177	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	114.360%
35) Dibromofluoromethane	5.397	113	42209	25.794	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	51.580%#
50) Toluene-d8	8.653	98	319470	53.139	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	106.280%
62) 4-Bromofluorobenzene	11.085	95	102925	50.347	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.700%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	99785	107.975	ug/l	90
20) Methylene Chloride	2.794	84	12486	6.946	ug/l #	88
25) 2-Butanone	4.574	43	8462	5.898	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081321\
Data File : VX023710.D
Acq On : 14 Aug 2021 00:04
Operator : JC/MD
Sample : M3371-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
JET-GROUT

Quant Time: Aug 14 05:33:00 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

