

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX081321\  
 Data File : VX023714.D  
 Acq On : 14 Aug 2021 01:39  
 Operator : JC/MD  
 Sample : M3394-11  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 41 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 SWCS-01-1015

Quant Time: Aug 14 05:34:41 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X072221W.M  
 Quant Title : SW846 8260  
 QLast Update : Fri Jul 23 06:05:38 2021  
 Response via : Initial Calibration

| Compound                    | R.T. QIon |       | Response | Conc     | Units | Dev(Min) |
|-----------------------------|-----------|-------|----------|----------|-------|----------|
| -----                       |           |       |          |          |       |          |
| Internal Standards          |           |       |          |          |       |          |
| 1) Pentafluorobenzene       | 5.562     | 168   | 148321   | 50.000   | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.769     | 114   | 253958   | 50.000   | ug/l  | 0.00     |
| 63) Chlorobenzene-d5        | 10.061    | 117   | 244041   | 50.000   | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.024    | 152   | 95634    | 50.000   | ug/l  | 0.00     |
| System Monitoring Compounds |           |       |          |          |       |          |
| 33) 1,2-Dichloroethane-d4   | 5.970     | 65    | 103668   | 56.031   | ug/l  | 0.00     |
| Spiked Amount               | 50.000    | Range | 78 - 117 | Recovery | =     | 112.060% |
| 35) Dibromofluoromethane    | 5.397     | 113   | 82859    | 52.183   | ug/l  | 0.00     |
| Spiked Amount               | 50.000    | Range | 75 - 124 | Recovery | =     | 104.360% |
| 50) Toluene-d8              | 8.653     | 98    | 320466   | 54.932   | ug/l  | 0.00     |
| Spiked Amount               | 50.000    | Range | 92 - 112 | Recovery | =     | 109.860% |
| 62) 4-Bromofluorobenzene    | 11.085    | 95    | 108020   | 54.453   | ug/l  | 0.00     |
| Spiked Amount               | 50.000    | Range | 83 - 123 | Recovery | =     | 108.900% |
| Target Compounds            |           |       |          |          |       |          |
|                             |           |       |          |          |       | Qvalue   |
| 16) Acetone                 | 2.385     | 43    | 29694    | 32.717   | ug/l  | 91       |
| 20) Methylene Chloride      | 2.794     | 84    | 13162    | 7.456    | ug/l  | 96       |
| 43) Isopropyl Acetate       | 6.354     | 43    | 21470    | 4.783    | ug/l  | 94       |
| -----                       |           |       |          |          |       |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX081321\  
Data File : VX023714.D  
Acq On : 14 Aug 2021 01:39  
Operator : JC/MD  
Sample : M3394-11  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 41 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
SWCS-01-1015

Quant Time: Aug 14 05:34:41 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X072221W.M  
Quant Title : SW846 8260  
QLast Update : Fri Jul 23 06:05:38 2021  
Response via : Initial Calibration

