

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081321\
 Data File : VX023720.D
 Acq On : 14 Aug 2021 04:00
 Operator : JC/MD
 Sample : M3397-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SWCS-07-0510

Quant Time: Aug 14 05:36:40 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	151519	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	258435	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	247339	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	96295	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	107396	56.821	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	113.640%
35) Dibromofluoromethane	5.397	113	84998	52.603	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.200%
50) Toluene-d8	8.653	98	326003	54.913	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	109.820%
62) 4-Bromofluorobenzene	11.085	95	108498	53.746	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	107.500%
Target Compounds						Qvalue
16) Acetone	2.386	43	28240	30.459	ug/l	88
20) Methylene Chloride	2.800	84	13124	7.277	ug/l	96
43) Isopropyl Acetate	6.348	43	21641	4.738	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081321\
Data File : VX023720.D
Acq On : 14 Aug 2021 04:00
Operator : JC/MD
Sample : M3397-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SWCS-07-0510

Quant Time: Aug 14 05:36:40 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

