

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081321\
 Data File : VX023721.D
 Acq On : 14 Aug 2021 04:24
 Operator : JC/MD
 Sample : M3397-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SWCS-07-1015

Quant Time: Aug 14 05:36:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 23 06:05:38 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	151625	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	263021	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	247492	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	94569	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	109021	57.641	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	115.280%
35) Dibromofluoromethane	5.397	113	84992	51.682	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.360%
50) Toluene-d8	8.653	98	333298	55.163	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	110.320%
62) 4-Bromofluorobenzene	11.085	95	107370	52.260	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	104.520%
Target Compounds						
						Qvalue
16) Acetone	2.392	43	29399	31.687	ug/l #	86
20) Methylene Chloride	2.794	84	13451	7.453	ug/l	88
43) Isopropyl Acetate	6.354	43	22309	4.799	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081321\
Data File : VX023721.D
Acq On : 14 Aug 2021 04:24
Operator : JC/MD
Sample : M3397-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SWCS-07-1015

Quant Time: Aug 14 05:36:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072221W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 23 06:05:38 2021
Response via : Initial Calibration

