

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081325\
 Data File : VX047317.D
 Acq On : 13 Aug 2025 13:18
 Operator : JC/MD
 Sample : Q2814-06RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TW-82H-SRE

Quant Time: Aug 14 03:28:02 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 22 03:59:51 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	190085	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.775	114	345739	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	321042	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	164920	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	141472	59.747	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	119.500%#
35) Dibromofluoromethane	5.404	113	122206	57.247	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	114.500%
50) Toluene-d8	8.653	98	331608	47.967	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.940%
62) 4-Bromofluorobenzene	11.079	95	165199	55.451	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	110.900%
Target Compounds						
16) Acetone	2.398	43	8598	9.064	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081325\
Data File : VX047317.D
Acq On : 13 Aug 2025 13:18
Operator : JC/MD
Sample : Q2814-06RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TW-82H-SRE

Quant Time: Aug 14 03:28:02 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 22 03:59:51 2025
Response via : Initial Calibration

