

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081325\
 Data File : VX047330.D
 Acq On : 13 Aug 2025 18:02
 Operator : JC/MD
 Sample : Q2815-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TW-88H-E

Quant Time: Aug 14 03:32:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 22 03:59:51 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	226623	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.775	114	419826	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	390050	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	198183	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	163301	57.847	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	115.700%
35) Dibromofluoromethane	5.403	113	143599	55.398	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	110.800%
50) Toluene-d8	8.653	98	374974	44.668	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	89.340%#
62) 4-Bromofluorobenzene	11.079	95	183003	50.587	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.180%
Target Compounds						
						Qvalue
16) Acetone	2.398	43	6333	5.600	ug/l	99
17) Carbon Disulfide	2.538	76	2398	0.291	ug/l #	88
19) Methyl tert-butyl Ether	3.136	73	4013	0.463	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081325\
Data File : VX047330.D
Acq On : 13 Aug 2025 18:02
Operator : JC/MD
Sample : Q2815-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TW-88H-E

Quant Time: Aug 14 03:32:56 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 22 03:59:51 2025
Response via : Initial Calibration

