

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081418\
 Data File : VX004031.D
 Acq On : 14 Aug 2018 15:31
 Operator : JC/MD
 Sample : J4429-01DL 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FL-0566DL

Quant Time: Aug 15 13:14:15 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 13:27:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	286646	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	444406	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	407135	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	252473	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	194138	48.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.14%	
35) Dibromofluoromethane	5.50	113	162431	48.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.68%	
50) Toluene-d8	8.71	98	639902	47.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.90%	
62) 4-Bromofluorobenzene	11.14	95	234600	47.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.94%	
Target Compounds						
						Qvalue
31) Cyclohexane	5.57	56	297084	53.21	ug/l	97
39) Methylcyclohexane	7.46	83	141276	24.63	ug/l	97
67) Ethyl Benzene	10.26	91	302171	18.14	ug/l	94
68) m/p-Xylenes	10.36	106	261693	40.13	ug/l	94
69) o-Xylene	10.70	106	4645	0.75	ug/l	91
73) Isopropylbenzene	11.02	105	45907	2.69	ug/l	98
78) n-propylbenzene	11.36	91	60540	3.17	ug/l	100
80) 1,3,5-Trimethylbenzene	11.51	105	142261	10.17	ug/l	99
84) 1,2,4-Trimethylbenzene	11.81	105	1034208	71.38	ug/l	100
86) p-Isopropyltoluene	12.07	119	17242	1.15	ug/l	92
95) Naphthalene	13.83	128	275050	16.37	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081418\
Data File : VX004031.D
Acq On : 14 Aug 2018 15:31
Operator : JC/MD
Sample : J4429-01DL 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FL-0566DL

Quant Time: Aug 15 13:14:15 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 13:27:25 2018
Response via : Initial Calibration

