

Data File : VX004034.D
Acq On : 14 Aug 2018 16:50
Operator : JC/MD
Sample : J4454-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CB-0818-S-TCLP-61A

Manual Integrations
APPROVED

MMDadoda
8/15/2018 4:49:54 PM

Quant Time: Aug 21 05:00:44 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 13:27:25 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	296841	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	475745	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	440763	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	252567	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	209601	50.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.24%	
35) Dibromofluoromethane	5.50	113	178037	50.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.00%	
50) Toluene-d8	8.71	98	677749	46.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.90%	
62) 4-Bromofluorobenzene	11.14	95	241871	46.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.38%	
Target Compounds						
16) Acetone	2.43	43	20222	9.300	ug/l	98
20) Methylene Chloride	2.86	84	8906	1.922	ug/l	94
43) Isopropyl Acetate	6.46	43	145598	16.755	ug/l	100
67) Ethyl Benzene	10.26	91	5344	0.296	ug/l	97
68) m/p-Xylenes	10.36	106	4166	0.590	ug/l	96
84) 1,2,4-Trimethylbenzene	11.81	105	8115	0.560	ug/l	97
95) Naphthalene	13.83	128	1895349	106.017	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data File : VX004034.D
Acq On : 14 Aug 2018 16:50
Operator : JC/MD
Sample : J4454-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CB-0818-S-TCLP-61A

Manual Integrations
APPROVED

MMDadoda
8/15/2018 4:49:54 PM

Quant Time: Aug 21 05:00:44 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 13:27:25 2018
Response via : Initial Calibration

