

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081418\
 Data File : VX004037.D
 Acq On : 14 Aug 2018 18:10
 Operator : JC/MD
 Sample : J4438-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-286-SRI3

Quant Time: Aug 15 15:42:05 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 13:27:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	280404	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	433674	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	410548	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	216911	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	197635	50.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.04%	
35) Dibromofluoromethane	5.51	113	167722	51.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.36%	
50) Toluene-d8	8.72	98	612016	46.51	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.02%	
62) 4-Bromofluorobenzene	11.14	95	211849	44.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.78%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	2.39	101	4907	1.55	ug/l	Qvalue 100
44) Trichloroethene	7.21	130	9868	2.45	ug/l	89
64) Tetrachloroethene	9.34	164	383724	98.62	ug/l	94
95) Naphthalene	13.83	128	1553	1.24	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081418\
Data File : VX004037.D
Acq On : 14 Aug 2018 18:10
Operator : JC/MD
Sample : J4438-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-286-SRI3

Quant Time: Aug 15 15:42:05 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 13:27:25 2018
Response via : Initial Calibration

