

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081418\
 Data File : VX004038.D
 Acq On : 14 Aug 2018 18:37
 Operator : JC/MD
 Sample : J4438-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RB1-SRI3

Manual Integrations
 APPROVED

MMDadoda
 8/15/2018 4:49:55 PM

Quant Time: Aug 15 15:43:20 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 13:27:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	278244	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	426891	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	400997	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	218169	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	195227	49.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.60%	
35) Dibromofluoromethane	5.51	113	160731	50.31	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.62%	
50) Toluene-d8	8.72	98	605633	46.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.50%	
62) 4-Bromofluorobenzene	11.14	95	207481	44.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.32%	
Target Compounds						
64) Tetrachloroethene	9.34	164	1626	0.43	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081418\
Data File : VX004038.D
Acq On : 14 Aug 2018 18:37
Operator : JC/MD
Sample : J4438-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RB1-SRI3

Manual Integrations
APPROVED

MMDadoda
8/15/2018 4:49:55 PM

Quant Time: Aug 15 15:43:20 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 13:27:25 2018
Response via : Initial Calibration

