

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081418\
 Data File : VX004040.D
 Acq On : 14 Aug 2018 19:30
 Operator : JC/MD
 Sample : J4438-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-291-SRI3

Quant Time: Aug 15 15:45:56 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 13:27:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	261905	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	414600	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	394859	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	209568	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	192106	52.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.12%	
35) Dibromofluoromethane	5.50	113	164337	52.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.92%	
50) Toluene-d8	8.72	98	593856	47.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.40%	
62) 4-Bromofluorobenzene	11.14	95	204630	44.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.70%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
9) 1,1,2-Trichlorotrifluoroet	2.39	101	4878	1.65	ug/l	98
12) 1,1-Dichloroethene	2.38	96	16201	5.88	ug/l	92
19) Methyl tert-butyl Ether	3.20	73	65429	6.74	ug/l	97
24) 1,1-Dichloroethane	3.70	63	8775	1.51	ug/l	98
27) cis-1,2-Dichloroethene	4.59	96	4034	1.15	ug/l	89
30) Chloroform	5.21	83	5250	0.92	ug/l	97
44) Trichloroethene	7.21	130	48065	12.50	ug/l	92
64) Tetrachloroethene	9.34	164	3261	0.87	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081418\
Data File : VX004040.D
Acq On : 14 Aug 2018 19:30
Operator : JC/MD
Sample : J4438-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-291-SRI3

Quant Time: Aug 15 15:45:56 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 13:27:25 2018
Response via : Initial Calibration

