

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081418\
 Data File : VX004041.D
 Acq On : 14 Aug 2018 19:57
 Operator : JC/MD
 Sample : J4438-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP1-SRI3

Quant Time: Aug 15 15:46:48 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 13:27:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	263916	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	420984	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	387760	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	214196	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	193403	52.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.02%	
35) Dibromofluoromethane	5.51	113	163485	51.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.78%	
50) Toluene-d8	8.72	98	597790	46.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.60%	
62) 4-Bromofluorobenzene	11.14	95	202248	43.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.30%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	2.39	101	76221	25.53	ug/l	99
19) Methyl tert-butyl Ether	3.20	73	28157	2.88	ug/l	99
30) Chloroform	5.22	83	21229	3.71	ug/l	95
38) Carbon Tetrachloride	5.79	117	5363	1.15	ug/l	100
44) Trichloroethene	7.21	130	20925	5.36	ug/l	97
64) Tetrachloroethene	9.34	164	49264	13.40	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081418\
Data File : VX004041.D
Acq On : 14 Aug 2018 19:57
Operator : JC/MD
Sample : J4438-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP1-SRI3

Quant Time: Aug 15 15:46:48 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 13:27:25 2018
Response via : Initial Calibration

