

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081418\
 Data File : VX004050.D
 Acq On : 15 Aug 2018 01:41
 Operator : JC/MD
 Sample : J4438-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-288-SRI3

Quant Time: Aug 15 15:50:31 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 13:27:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	260076	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	411991	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	387769	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	210917	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	190838	52.08	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.16%	
35) Dibromofluoromethane	5.51	113	162230	52.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.22%	
50) Toluene-d8	8.72	98	599875	47.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.96%	
62) 4-Bromofluorobenzene	11.14	95	201629	44.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.94%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	2.39	101	108282	36.81	ug/l	Qvalue 100
19) Methyl tert-butyl Ether	3.20	73	23176	2.40	ug/l	99
30) Chloroform	5.22	83	22850	4.05	ug/l	97
44) Trichloroethene	7.21	130	21340	5.58	ug/l	89
64) Tetrachloroethene	9.34	164	31717	8.63	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081418\
Data File : VX004050.D
Acq On : 15 Aug 2018 01:41
Operator : JC/MD
Sample : J4438-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-288-SRI3

Quant Time: Aug 15 15:50:31 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 13:27:25 2018
Response via : Initial Calibration

