

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081418\
 Data File : VX004051.D
 Acq On : 15 Aug 2018 02:08
 Operator : JC/MD
 Sample : J4438-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-297-SRI3

Quant Time: Aug 15 15:50:57 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 13:27:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	277328	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	439475	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	397617	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	223066	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	202273	51.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.54%	
35) Dibromofluoromethane	5.51	113	171591	52.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.34%	
50) Toluene-d8	8.72	98	620766	46.55	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.10%	
62) 4-Bromofluorobenzene	11.14	95	211001	43.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.24%	
Target Compounds						
44) Trichloroethene	7.21	130	66454	16.30	ug/l	96
64) Tetrachloroethene	9.34	164	2020	0.54	ug/l #	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081418\
Data File : VX004051.D
Acq On : 15 Aug 2018 02:08
Operator : JC/MD
Sample : J4438-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-297-SRI3

Quant Time: Aug 15 15:50:57 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 13:27:25 2018
Response via : Initial Calibration

