

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081418\
 Data File : VX004053.D
 Acq On : 15 Aug 2018 03:01
 Operator : JC/MD
 Sample : J4438-12
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RB3-SRI3

Quant Time: Aug 15 15:52:05 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 13:27:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	272552	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	429249	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	395196	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	229850	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	200556	52.23	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.46%	
35) Dibromofluoromethane	5.51	113	167055	52.00	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.00%	
50) Toluene-d8	8.72	98	617987	47.45	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.90%	
62) 4-Bromofluorobenzene	11.14	95	213508	45.19	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.38%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081418\
Data File : VX004053.D
Acq On : 15 Aug 2018 03:01
Operator : JC/MD
Sample : J4438-12
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RB3-SRI3

Quant Time: Aug 15 15:52:05 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 13:27:25 2018
Response via : Initial Calibration

