

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081418\
 Data File : VX004060.D
 Acq On : 15 Aug 2018 06:07
 Operator : JC/MD
 Sample : J4439-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP2-SRI3

Quant Time: Aug 15 15:57:00 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 13:27:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	253555	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	403404	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	387008	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	211634	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	194325	54.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.80%	
35) Dibromofluoromethane	5.50	113	163414	54.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.26%	
50) Toluene-d8	8.72	98	590468	48.24	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.48%	
62) 4-Bromofluorobenzene	11.14	95	198449	44.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.40%	
Target Compounds						
44) Trichloroethene	7.21	130	13790	3.69	ug/l	96
64) Tetrachloroethene	9.34	164	36241	9.88	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081418\
Data File : VX004060.D
Acq On : 15 Aug 2018 06:07
Operator : JC/MD
Sample : J4439-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP2-SRI3

Quant Time: Aug 15 15:57:00 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 13:27:25 2018
Response via : Initial Calibration

