

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081418\
 Data File : VX004065.D
 Acq On : 15 Aug 2018 08:19
 Operator : JC/MD
 Sample : J4439-16
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-287-SRI3

Quant Time: Aug 15 10:43:04 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 13:27:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	253429	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	400031	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	384705	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	202752	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	191430	53.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.22%	
35) Dibromofluoromethane	5.51	113	156551	52.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.58%	
50) Toluene-d8	8.72	98	574544	47.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.66%	
62) 4-Bromofluorobenzene	11.14	95	193849	44.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.06%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	2.39	101	10978	3.83	ug/l	99
12) 1,1-Dichloroethene	2.38	96	3336	1.25	ug/l	94
24) 1,1-Dichloroethane	3.70	63	11509	2.04	ug/l	98
44) Trichloroethene	7.21	130	9736	2.62	ug/l	92
64) Tetrachloroethene	9.34	164	3564	0.98	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081418\
Data File : VX004065.D
Acq On : 15 Aug 2018 08:19
Operator : JC/MD
Sample : J4439-16
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-287-SRI3

Quant Time: Aug 15 10:43:04 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 13:27:25 2018
Response via : Initial Calibration

