

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081418\
 Data File : VX004057.D
 Acq On : 15 Aug 2018 04:47
 Operator : JC/MD
 Sample : J4439-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RB5-SRI3

Quant Time: Aug 15 11:08:05 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 13:27:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	254302	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	408163	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	372406	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	207093	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	191977	53.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.16%	
35) Dibromofluoromethane	5.51	113	162920	53.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.66%	
50) Toluene-d8	8.72	98	578735	46.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.46%	
62) 4-Bromofluorobenzene	11.14	95	194071	43.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.40%	


Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081418\
Data File : VX004057.D
Acq On : 15 Aug 2018 04:47
Operator : JC/MD
Sample : J4439-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleID :
RB5-SRI3

Quant Time: Aug 15 11:08:05 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 13:27:25 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.00 ug/l

RT: 5.67 min Scan# 833

Delta R.T. -0.00 min

Lab File: VX004057.D

Acq: 15 Aug 2018 04:47

Instrument :

MSVOA_X

ClientSampleId :


RB5-SRI3

Tot Ion:168 Resp: 254302

Ion Ratio Lower Upper

168 100

99 47.6 40.1 60.1



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.67

100000

80000

60000



40000

20000

0

Time-->

5.50 5.60 5.70 5.80 5.90



#33

1,2-Dichloroethane-d4

Concen: 53.58 ug/l

RT: 6.07 min Scan# 899

Delta R.T. -0.00 min

Lab File: VX004057.D


Acq: 15 Aug 2018 04:47

Tgt Ion: 65 Resp: 191977

Ion Ratio Lower Upper

65 100

67 52.7 0.0 105.8



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0

6.07

80000

60000


40000


20000

0

Time-->

5.90 6.00 6.10 6.20









#34
 1,4-Difluorobenzene
 Concen: 50.00 ug/l
 RT: 6.86 min Scan# 1029
 Delta R.T. -0.01 min
 Lab File: VX004057.D
 Acq: 15 Aug 2018 04:47




Instrument :
 MSVOA_X
 ClientSampleId :
 RB5-SRI3


Tot Ion:114 Resp: 408163
 Ion Ratio Lower Upper
 114 100
 63 21.1 0.0 39.0
 88 15.5 0.0 32.0



#35
 Dibromofluoromethane
 Concen: 53.33 ug/l
 RT: 5.51 min Scan# 807
 Delta R.T. 0.01 min
 Lab File: VX004057.D
 Acq: 15 Aug 2018 04:47

Tgt Ion:113 Resp: 162920
 Ion Ratio Lower Upper
 113 100
 111 102.7 82.0 123.0
 192 23.7 19.0 28.4









#50
Toluene-d8
Concen: 46.73 ug/l
RT: 8.72 min Scan# 1334
Delta R.T. -0.00 min
Lab File: VX004057.D
Acq: 15 Aug 2018 04:47




Instrument :
MSVOA_X
ClientSampleId :
RB5-SRI3


Tot Ion: 98 Resp: 578735
Ion Ratio Lower Upper
98 100
100 65.7 53.3 79.9



#62
4-Bromofluorobenzene
Concen: 43.20 ug/l
RT: 11.14 min Scan# 1731
Delta R.T. -0.00 min
Lab File: VX004057.D
Acq: 15 Aug 2018 04:47

Tgt Ion: 95 Resp: 194071
Ion Ratio Lower Upper
95 100
174 93.0 0.0 182.8
176 89.7 0.0 179.0









#63
Chlorobenzene-d5
Concen: 50.00 ug/l
RT: 10.12 min Scan# 1563
Delta R.T. -0.00 min
Lab File: VX004057.D
Acq: 15 Aug 2018 04:47

Instrument :
MSVOA_X
ClientSampleId :
RB5-SRI3

Tot Ion:117 Resp: 372406
Ion Ratio Lower Upper
117 100
82 55.4 45.4 68.0
119 33.4 26.6 40.0




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): VX0
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.00 ug/l
RT: 12.08 min Scan# 1885
Delta R.T. -0.00 min
Lab File: VX004057.D
Acq: 15 Aug 2018 04:47

Tgt Ion:152 Resp: 207093
Ion Ratio Lower Upper
152 100
115 54.3 39.1 117.3
150 154.6 0.0 349.2



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

