

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081420\
 Data File : VX017904.D
 Acq On : 14 Aug 2020 17:37
 Operator : JC/SP
 Sample : L3697-09
 Misc : 25.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BG0N7

Quant Time: Aug 17 03:49:19 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\SOMXTR080720WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Aug 17 01:26:51 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	6.83	114	295687	5.00	ug/L	0.00
28) Chlorobenzene-d5	10.10	117	275689	5.00	ug/L	0.00
61) 1,4-Dichlorobenzene-d4	12.06	152	128580	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.39	65	90168	4.41	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	88.20%
7) Chloroethane-d5	1.69	69	75178	4.63	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	92.60%
11) 1,1-Dichloroethene-d2	2.34	63	137507	3.27	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	65.40%
20) 2-Butanone-d5	4.53	46	238026	43.87	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	87.74%
24) Chloroform-d	5.14	84	161993	3.91	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	78.20%
26) 1,2-Dichloroethane-d4	6.03	65	106888	4.69	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	93.80%
32) Benzene-d6	6.05	84	353215	4.67	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	93.40%
36) 1,2-Dichloropropane-d6	7.37	67	107683	4.60	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.00%
41) Toluene-d8	8.70	98	291196	4.04	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	80.80%
45) trans-1,3-Dichloropropene-	8.99	79	41151	4.06	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	81.20%
48) 2-Hexanone-d5	9.43	63	196362	46.63	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	93.26%
59) 1,1,2,2-Tetrachloroethane-	11.23	84	75634	4.25	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	85.00%
65) 1,2-Dichlorobenzene-d4	12.36	152	123995	5.01	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	100.20%

Target Compounds					Ovalue
25) Chloroform	5.18	83	17998	0.364 ug/L	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081420\
Data File : VX017904.D
Acq On : 14 Aug 2020 17:37
Operator : JC/SP
Sample : L3697-09
Misc : 25.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BG0N7

Quant Time: Aug 17 03:49:19 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\SOMXTR080720WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Aug 17 01:26:51 2020
Response via : Initial Calibration

