

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081423\
 Data File : VX036955.D
 Acq On : 14 Aug 2023 10:48
 Operator : JC/MD
 Sample : VX0814MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0814MBL01

Quant Time: Aug 15 01:27:14 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	140478	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	255906	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	242047	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	107951	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	97401	46.859	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	93.720%
35) Dibromofluoromethane	5.379	113	82570	46.943	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.880%
50) Toluene-d8	8.647	98	307591	49.312	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.620%
62) 4-Bromofluorobenzene	11.079	95	118242	50.244	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.480%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081423\
Data File : VX036955.D
Acq On : 14 Aug 2023 10:48
Operator : JC/MD
Sample : VX0814MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0814MBL01

Quant Time: Aug 15 01:27:14 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

