

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081423\
 Data File : VX036963.D
 Acq On : 14 Aug 2023 14:36
 Operator : JC/MD
 Sample : 03997-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TB-20230811

Quant Time: Aug 15 01:29:11 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	125388	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	232403	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	218911	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	100294	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	88226	47.553	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.100%
35) Dibromofluoromethane	5.391	113	74171	46.432	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.860%
50) Toluene-d8	8.647	98	275229	48.586	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.180%
62) 4-Bromofluorobenzene	11.079	95	107872	50.473	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.940%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	2793	3.628	ug/l	97
20) Methylene Chloride	2.794	84	3080	1.799	ug/l	86
94) Hexachlorobutadiene	13.725	225	207	0.240	ug/l	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081423\
Data File : VX036963.D
Acq On : 14 Aug 2023 14:36
Operator : JC/MD
Sample : 03997-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TB-20230811

Quant Time: Aug 15 01:29:11 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

