

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081423\
 Data File : VX036966.D
 Acq On : 14 Aug 2023 15:48
 Operator : JC/MD
 Sample : 03997-01RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TB-20230811RE

Quant Time: Aug 15 01:29:37 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	121250	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	221954	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	204189	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	85159	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	85748	47.795	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.600%
35) Dibromofluoromethane	5.385	113	72292	47.386	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.780%
50) Toluene-d8	8.647	98	263774	48.756	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.520%
62) 4-Bromofluorobenzene	11.079	95	98679	48.345	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.700%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	2928	3.933	ug/l #	80
20) Methylene Chloride	2.788	84	1671	1.009	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081423\
Data File : VX036966.D
Acq On : 14 Aug 2023 15:48
Operator : JC/MD
Sample : 03997-01RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TB-20230811RE

Quant Time: Aug 15 01:29:37 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

