

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081423\
 Data File : VX036971.D
 Acq On : 14 Aug 2023 17:48
 Operator : JC/MD
 Sample : 03986-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1770

Quant Time: Aug 15 01:30:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	112227	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	208045	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	199213	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	92500	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	81126	48.854	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.700%
35) Dibromofluoromethane	5.385	113	67322	47.079	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.160%
50) Toluene-d8	8.647	98	251837	49.662	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.320%
62) 4-Bromofluorobenzene	11.079	95	99982	52.258	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	104.520%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	145416	211.032	ug/l	93
25) 2-Butanone	4.562	43	42346	37.877	ug/l #	89
40) Benzene	6.050	78	6628	1.069	ug/l #	89
52) Toluene	8.720	92	9826	2.529	ug/l	97
69) o-Xylene	10.640	106	4597	1.539	ug/l	98
84) 1,2,4-Trimethylbenzene	11.750	105	6774	1.058	ug/l	98
95) Naphthalene	13.774	128	22177	3.315	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081423\
Data File : VX036971.D
Acq On : 14 Aug 2023 17:48
Operator : JC/MD
Sample : 03986-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1770

Quant Time: Aug 15 01:30:35 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

