

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081423\
 Data File : VX036974.D
 Acq On : 14 Aug 2023 19:00
 Operator : JC/MD
 Sample : 03932-03 10X
 Misc : 7.14g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 NEVINS-SB-01-Z30-32

Quant Time: Aug 15 01:31:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	121704	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	225521	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	224218	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	113322	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	84972	47.186	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.380%
35) Dibromofluoromethane	5.385	113	71685	46.245	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.500%
50) Toluene-d8	8.647	98	275399	50.100	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.200%
62) 4-Bromofluorobenzene	11.079	95	116408	56.129	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	112.260%
Target Compounds						
					Qvalue	
20) Methylene Chloride	2.788	84	1012	0.609	ug/l	89
39) Methylcyclohexane	7.379	83	6594	2.295	ug/l	90
40) Benzene	6.038	78	151951	22.605	ug/l	99
52) Toluene	8.720	92	1650	0.392	ug/l	98
67) Ethyl Benzene	10.195	91	450832	49.932	ug/l	99
68) m/p-Xylenes	10.305	106	17294	5.134	ug/l	97
69) o-Xylene	10.640	106	18799	5.591	ug/l	95
73) Isopropylbenzene	10.964	105	38872	4.145	ug/l	98
78) n-propylbenzene	11.305	91	21673	1.998	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	38525	4.910	ug/l	98
84) 1,2,4-Trimethylbenzene	11.750	105	186606	23.782	ug/l	96
86) p-Isopropyltoluene	12.006	119	16853	2.131	ug/l #	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081423\
Data File : VX036974.D
Acq On : 14 Aug 2023 19:00
Operator : JC/MD
Sample : 03932-03 10X
Misc : 7.14g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
NEVINS-SB-01-Z30-32

Quant Time: Aug 15 01:31:12 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

