

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081423\
 Data File : VX036954.D
 Acq On : 14 Aug 2023 10:17
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 15 01:26:53 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	82	0.00
2 T	Dichlorodifluoromethane	0.661	0.694	-5.0	91	0.00
3 P	Chloromethane	0.601	0.556	7.5	80	0.00
4 C	Vinyl Chloride	0.612	0.623	-1.8#	84	0.00
5 T	Bromomethane	0.365	0.412	-12.9	96	-0.02
6 T	Chloroethane	0.297	0.396	-33.3#	105	-0.01
7 T	Trichlorofluoromethane	0.945	1.134	-20.0	100	0.00
8 T	Diethyl Ether	0.365	0.397	-8.8	95	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.567	0.640	-12.9	93	0.00
10 T	Methyl Iodide	0.589	0.770	-30.7#	101	0.00
11 T	Tert butyl alcohol	0.175	0.163	6.9	77	-0.03
12 CM	1,1-Dichloroethene	0.560	0.591	-5.5#	89	0.00
13 T	Acrolein	0.100	0.098	2.0	84	0.00
14 T	Allyl chloride	1.102	0.988	10.3	75	0.00
15 T	Acrylonitrile	0.352	0.344	2.3	79	0.00
16 T	Acetone	0.307	0.312	-1.6	86	-0.01
17 T	Carbon Disulfide	1.545	1.433	7.2	78	0.00
18 T	Methyl Acetate	1.291	1.228	4.9	77	0.00
19 T	Methyl tert-butyl Ether	2.104	2.282	-8.5	89	0.00
20 T	Methylene Chloride	0.683	0.674	1.3	86	0.00
21 T	trans-1,2-Dichloroethene	0.626	0.654	-4.5	88	0.00
22 T	Diisopropyl ether	2.093	1.993	4.8	79	0.00
23 T	Vinyl Acetate	1.606	1.538	4.2	77	0.00
24 P	1,1-Dichloroethane	1.174	1.213	-3.3	85	0.00
25 T	2-Butanone	0.498	0.476	4.4	78	-0.01
26 T	2,2-Dichloropropane	1.106	1.197	-8.2	91	0.00
27 T	cis-1,2-Dichloroethene	0.743	0.790	-6.3	90	0.00
28 T	Bromochloromethane	0.568	0.524	7.7	75	0.00
29 T	Tetrahydrofuran	0.321	0.293	8.7	74	-0.01
30 C	Chloroform	1.201	1.323	-10.2#	92	-0.01
31 T	Cyclohexane	1.024	0.965	5.8	78	0.00
32 T	1,1,1-Trichloroethane	1.100	1.224	-11.3	92	0.00
33 S	1,2-Dichloroethane-d4	0.740	0.692	6.5	79	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	79	0.00
35 S	Dibromofluoromethane	0.344	0.349	-1.5	83	0.00
36 T	1,1-Dichloropropene	0.512	0.562	-9.8	89	-0.01
37 T	Ethyl Acetate	0.580	0.556	4.1	75	0.00
38 T	Carbon Tetrachloride	0.545	0.639	-17.2	94	0.00
39 T	Methylcyclohexane	0.637	0.616	3.3	78	0.00
40 TM	Benzene	1.490	1.609	-8.0	86	0.00
41 T	Methacrylonitrile	0.310	0.303	2.3	76	0.00
42 TM	1,2-Dichloroethane	0.545	0.630	-15.6	91	0.00
43 T	Isopropyl Acetate	0.998	0.956	4.2	77	0.00
44 TM	Trichloroethene	0.390	0.450	-15.4	92	0.00
45 C	1,2-Dichloropropane	0.398	0.416	-4.5#	83	0.00
46 T	Dibromomethane	0.278	0.314	-12.9	89	0.00
47 T	Bromodichloromethane	0.572	0.642	-12.2	89	0.00
48 T	Methyl methacrylate	0.464	0.456	1.7	77	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081423\
 Data File : VX036954.D
 Acq On : 14 Aug 2023 10:17
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 15 01:26:53 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 21 12:14:12 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.010	-11.1	81	-0.04
50 S	Toluene-d8	1.219	1.136	6.8	74	0.00
51 T	4-Methyl-2-Pentanone	0.552	0.549	0.5	76	0.00
52 CM	Toluene	0.934	1.057	-13.2#	89	0.00
53 T	t-1,3-Dichloropropene	0.646	0.709	-9.8	85	0.00
54 T	cis-1,3-Dichloropropene	0.682	0.739	-8.4	85	0.00
55 T	1,1,2-Trichloroethane	0.393	0.440	-12.0	89	0.00
56 T	Ethyl methacrylate	0.639	0.699	-9.4	84	0.00
57 T	1,3-Dichloropropane	0.654	0.736	-12.5	87	0.00
58 T	2-Chloroethyl Vinyl ether	0.315	0.313	0.6	78	0.00
59 T	2-Hexanone	0.410	0.420	-2.4	79	0.00
60 T	Dibromochloromethane	0.431	0.507	-17.6	91	0.00
61 T	1,2-Dibromoethane	0.407	0.473	-16.2	91	0.00
62 S	4-Bromofluorobenzene	0.460	0.476	-3.5	79	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	78	0.00
64 T	Tetrachloroethene	0.373	0.460	-23.3	97	0.00
65 PM	Chlorobenzene	1.102	1.271	-15.3	90	0.00
66 T	1,1,1,2-Tetrachloroethane	0.426	0.499	-17.1	92	0.00
67 C	Ethyl Benzene	2.013	2.264	-12.5#	88	0.00
68 T	m/p-Xylenes	0.751	0.866	-15.3	89	0.00
69 T	o-Xylene	0.750	0.861	-14.8	88	0.00
70 T	Styrene	1.237	1.442	-16.6	89	0.00
71 P	Bromoform	0.345	0.403	-16.8	89	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	82	0.00
73 T	Isopropylbenzene	4.138	4.357	-5.3	88	0.00
74 T	N-amyl acetate	1.930	1.736	10.1	75	0.00
75 P	1,1,2,2-Tetrachloroethane	1.398	1.393	0.4	85	0.00
76 T	1,2,3-Trichloropropane	1.176	1.244	-5.8	87	0.00
77 T	Bromobenzene	0.942	1.051	-11.6	91	0.00
78 T	n-propylbenzene	4.786	4.978	-4.0	85	0.00
79 T	2-Chlorotoluene	2.900	3.044	-5.0	88	0.00
80 T	1,3,5-Trimethylbenzene	3.462	3.669	-6.0	87	0.00
81 T	trans-1,4-Dichloro-2-butene	0.519	0.498	4.0	77	0.00
82 T	4-Chlorotoluene	3.304	3.540	-7.1	89	0.00
83 T	tert-Butylbenzene	3.401	3.491	-2.6	85	0.00
84 T	1,2,4-Trimethylbenzene	3.462	3.684	-6.4	88	0.00
85 T	sec-Butylbenzene	4.241	4.251	-0.2	82	0.00
86 T	p-Isopropyltoluene	3.490	3.616	-3.6	84	0.00
87 T	1,3-Dichlorobenzene	1.745	1.947	-11.6	91	0.00
88 T	1,4-Dichlorobenzene	1.732	1.939	-12.0	92	0.00
89 T	n-Butylbenzene	3.124	3.102	0.7	80	0.00
90 T	Hexachloroethane	0.707	0.667	5.7	77	0.00
91 T	1,2-Dichlorobenzene	1.684	1.891	-12.3	93	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.325	0.333	-2.5	87	0.00
93 T	1,2,4-Trichlorobenzene	1.038	1.139	-9.7	89	0.00
94 T	Hexachlorobutadiene	0.429	0.460	-7.2	90	0.00
95 T	Naphthalene	3.616	4.015	-11.0	91	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081423\
Data File : VX036954.D
Acq On : 14 Aug 2023 10:17
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Aug 15 01:26:53 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072123W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 21 12:14:12 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.023	1.116	-9.1	89	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6