

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081424\
 Data File : VX043030.D
 Acq On : 14 Aug 2024 14:37
 Operator : JC/MD
 Sample : VX0814WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0814WBL01

Quant Time: Aug 15 04:32:09 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081424W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 15 04:22:11 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	93153	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	162891	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	142295	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	64033	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	55194	45.092	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	90.180%
35) Dibromofluoromethane	5.385	113	49410	49.032	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.060%
50) Toluene-d8	8.647	98	196155	50.426	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.860%
62) 4-Bromofluorobenzene	11.079	95	67251	47.679	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	95.360%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081424\
Data File : VX043030.D
Acq On : 14 Aug 2024 14:37
Operator : JC/MD
Sample : VX0814WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0814WBL01

Quant Time: Aug 15 04:32:09 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081424W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 15 04:22:11 2024
Response via : Initial Calibration

