

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081424\
 Data File : VX043034.D
 Acq On : 14 Aug 2024 16:19
 Operator : JC/MD
 Sample : P3542-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Storage-Blank-SOIL-REF-6

Quant Time: Aug 15 04:35:01 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081424W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 15 04:22:11 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	81704	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	142385	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	122869	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	49103	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	48917	45.564	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	91.120%
35) Dibromofluoromethane	5.385	113	43745	49.663	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.320%
50) Toluene-d8	8.647	98	172121	50.620	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	101.240%
62) 4-Bromofluorobenzene	11.079	95	55030	44.633	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	89.260%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081424\
Data File : VX043034.D
Acq On : 14 Aug 2024 16:19
Operator : JC/MD
Sample : P3542-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Storage-Blank-SOIL-REF-6

Quant Time: Aug 15 04:35:01 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081424W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 15 04:22:11 2024
Response via : Initial Calibration

