

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081424\
 Data File : VX043045.D
 Acq On : 14 Aug 2024 20:37
 Operator : JC/MD
 Sample : P3574-02 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-W-06-202408

Quant Time: Aug 15 04:37:12 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081424W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 15 04:22:11 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	98572	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	172397	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	146392	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	65918	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	59713	46.102	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	92.200%
35) Dibromofluoromethane	5.385	113	52200	48.945	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.880%
50) Toluene-d8	8.647	98	202360	49.153	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.300%
62) 4-Bromofluorobenzene	11.079	95	68948	46.187	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	92.380%
Target Compounds						
					Qvalue	
20) Methylene Chloride	2.788	84	2297	1.546	ug/l #	88
30) Chloroform	5.099	83	3164	1.435	ug/l	94
64) Tetrachloroethene	9.275	164	2266	2.398	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081424\
Data File : VX043045.D
Acq On : 14 Aug 2024 20:37
Operator : JC/MD
Sample : P3574-02 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-W-06-202408

Quant Time: Aug 15 04:37:12 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081424W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 15 04:22:11 2024
Response via : Initial Calibration

