

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081424\
 Data File : VX043043.D
 Acq On : 14 Aug 2024 19:50
 Operator : JC/MD
 Sample : P3542-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Storage-Blank-SAMPLE-MANAGEMENT

Quant Time: Aug 15 04:36:12 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081424W.M
 Quant Title : SW846 8260
 QLast Update : Thu Aug 15 04:22:11 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	83120	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	146001	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	126906	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	52967	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	50431	46.174	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	92.340%
35) Dibromofluoromethane	5.385	113	44618	49.399	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.800%
50) Toluene-d8	8.647	98	174752	50.121	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.240%
62) 4-Bromofluorobenzene	11.079	95	57694	45.635	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	91.280%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081424\
Data File : VX043043.D
Acq On : 14 Aug 2024 19:50
Operator : JC/MD
Sample : P3542-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Storage-Blank-SAMPLE-MANAGEMENT

Quant Time: Aug 15 04:36:12 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X081424W.M
Quant Title : SW846 8260
QLast Update : Thu Aug 15 04:22:11 2024
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.550 min Scan# 71

Delta R.T. -0.000 min

Lab File: VX043043.D

Acq: 14 Aug 2024 19:50

Instrument :

MSVOA_X

ClientSampleId :





Storage-Blank-SAMPLE-MANAGEMENT

Tgt Ion:168 Resp: 83120

Ion Ratio Lower Upper

168 100

99 57.5 56.6 85.0



#33

1,2-Dichloroethane-d4

Concen: 46.174 ug/l

RT: 5.958 min Scan# 799

Delta R.T. 0.000 min

Lab File: VX043043.D




Acq: 14 Aug 2024 19:50


Tgt Ion: 65 Resp: 50431

Ion Ratio Lower Upper

65 100

67 56.1 0.0 100.2









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.763 min Scan# 911
Delta R.T. -0.000 min
Lab File: VX043043.D
Acq: 14 Aug 2024 19:50




Instrument :
MSVOA_X
ClientSampleId :
Storage-Blank-SAMPLE-MANAGEMENT


Tgt Ion:114 Resp: 146001
Ion Ratio Lower Upper
114 100
63 20.1 0.0 45.0
88 15.1 0.0 36.6



#35
Dibromofluoromethane
Concen: 49.399 ug/l
RT: 5.385 min Scan# 705
Delta R.T. 0.000 min
Lab File: VX043043.D
Acq: 14 Aug 2024 19:50

Tgt Ion:113 Resp: 44618
Ion Ratio Lower Upper
113 100
111 101.9 82.9 124.3
192 18.5 13.3 19.9









#50
Toluene-d8
Concen: 50.121 ug/l
RT: 8.647 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX043043.D
Acq: 14 Aug 2024 19:50




Instrument : MSVOA_X
ClientSampleId :
Storage-Blank-SAMPLE-MANAGEMENT


Tgt Ion: 98 Resp: 174752
Ion Ratio Lower Upper
98 100
100 66.3 53.1 79.7



#62
4-Bromofluorobenzene
Concen: 45.635 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. 0.000 min
Lab File: VX043043.D
Acq: 14 Aug 2024 19:50

Tgt Ion: 95 Resp: 57694
Ion Ratio Lower Upper
95 100
174 76.4 0.0 135.6
176 74.0 0.0 131.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX043043.D
Acq: 14 Aug 2024 19:50

Instrument : MSVOA_X
ClientSampleId :
Storage-Blank-SAMPLE-MANAGEMENT

Tgt Ion:117 Resp: 126906
Ion Ratio Lower Upper
117 100
82 52.3 46.2 69.4
119 32.5 25.5 38.3



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. -0.000 min
Lab File: VX043043.D
Acq: 14 Aug 2024 19:50

Tgt Ion:152 Resp: 52967
Ion Ratio Lower Upper
152 100
115 59.6 43.3 129.9
150 160.9 0.0 345.0

