

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081425\
 Data File : VX047339.D
 Acq On : 14 Aug 2025 12:47
 Operator : JC/MD
 Sample : Q2814-21
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB

Quant Time: Aug 18 01:28:02 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 22 03:59:51 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	228684	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	436084	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	425164	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	212249	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	162693	57.112	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 114.220%	
35) Dibromofluoromethane	5.391	113	135473	50.314	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.620%	
50) Toluene-d8	8.653	98	479102	54.945	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 109.880%	
62) 4-Bromofluorobenzene	11.079	95	194775	51.834	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 103.660%	
Target Compounds						
16) Acetone	2.398	43	4133	3.622	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081425\
Data File : VX047339.D
Acq On : 14 Aug 2025 12:47
Operator : JC/MD
Sample : Q2814-21
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB

Quant Time: Aug 18 01:28:02 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 22 03:59:51 2025
Response via : Initial Calibration

