

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081425\
 Data File : VX047341.D
 Acq On : 14 Aug 2025 13:58
 Operator : JC/MD
 Sample : Q2814-17
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TW-518R-S

Quant Time: Aug 18 01:29:12 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 22 03:59:51 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	230111	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	440338	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	435075	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	217218	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.971	65	159597	55.677	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	111.360%
35) Dibromofluoromethane	5.404	113	134982	49.648	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.300%
50) Toluene-d8	8.653	98	483933	54.963	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	109.920%
62) 4-Bromofluorobenzene	11.079	95	202666	53.412	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	106.820%
Target Compounds						
						Qvalue
16) Acetone	2.398	43	8672	7.552	ug/l	92
17) Carbon Disulfide	2.532	76	7883	0.942	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081425\
Data File : VX047341.D
Acq On : 14 Aug 2025 13:58
Operator : JC/MD
Sample : Q2814-17
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TW-518R-S

Quant Time: Aug 18 01:29:12 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X072125W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 22 03:59:51 2025
Response via : Initial Calibration

