

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081518\
 Data File : VX004077.D
 Acq On : 15 Aug 2018 17:21
 Operator : JC/MD
 Sample : J4428-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LF1-TB-2018-2

Quant Time: Aug 16 16:52:42 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 13:27:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	257912	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	399668	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	356633	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	189575	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	186372	51.29	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.58%	
35) Dibromofluoromethane	5.51	113	157761	52.74	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.48%	
50) Toluene-d8	8.72	98	564349	46.53	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.06%	
62) 4-Bromofluorobenzene	11.14	95	190127	43.22	ug/l	0.00
Spiked Amount	50.000		Recovery	=	86.44%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081518\
Data File : VX004077.D
Acq On : 15 Aug 2018 17:21
Operator : JC/MD
Sample : J4428-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LF1-TB-2018-2

Quant Time: Aug 16 16:52:42 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 13:27:25 2018
Response via : Initial Calibration

