

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081518\
 Data File : VX004079.D
 Acq On : 15 Aug 2018 18:14
 Operator : JC/MD
 Sample : J4473-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-01-180813

Quant Time: Aug 16 16:54:11 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 13:27:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	276286	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	419951	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	393829	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	201737	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	195820	50.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.60%	
35) Dibromofluoromethane	5.51	113	160648	51.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.22%	
50) Toluene-d8	8.72	98	595800	46.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.50%	
62) 4-Bromofluorobenzene	11.14	95	200025	43.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.56%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.44	43	10660	5.27	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081518\
Data File : VX004079.D
Acq On : 15 Aug 2018 18:14
Operator : JC/MD
Sample : J4473-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-01-180813

Quant Time: Aug 16 16:54:11 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 13:27:25 2018
Response via : Initial Calibration

