

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081518\
 Data File : VX004087.D
 Acq On : 15 Aug 2018 21:48
 Operator : JC/MD
 Sample : J4439-20
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-293-SRI3

Quant Time: Aug 16 07:02:40 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 13:27:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	235113	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	374001	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	364611	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	187888	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	181407	54.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.52%	
35) Dibromofluoromethane	5.50	113	153074	54.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.38%	
50) Toluene-d8	8.72	98	541161	47.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.36%	
62) 4-Bromofluorobenzene	11.14	95	181280	44.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.08%	
Target Compounds						
9) 1,1,2-Trichlorotrifluoroet	2.39	101	2347	0.88	ug/l	99
16) Acetone	2.44	43	3226	1.87	ug/l #	87
44) Trichloroethene	7.22	130	3418	0.99	ug/l	82
64) Tetrachloroethene	9.34	164	4071	1.18	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081518\
Data File : VX004087.D
Acq On : 15 Aug 2018 21:48
Operator : JC/MD
Sample : J4439-20
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-293-SRI3

Quant Time: Aug 16 07:02:40 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 13:27:25 2018
Response via : Initial Calibration

