

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081518\
 Data File : VX004088.D
 Acq On : 15 Aug 2018 22:14
 Operator : JC/MD
 Sample : J4439-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-295-SRI3

Quant Time: Aug 16 17:00:04 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 13:27:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	239726	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	358343	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	344562	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	182290	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	171085	50.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.30%	
35) Dibromofluoromethane	5.51	113	141380	52.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.44%	
50) Toluene-d8	8.72	98	510532	46.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.90%	
62) 4-Bromofluorobenzene	11.14	95	169577	43.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.00%	
Target Compounds						
7) Trichlorofluoromethane	1.93	101	2518	0.58	ug/l	89
9) 1,1,2-Trichlorotrifluoroet	2.39	101	2154	0.79	ug/l	98
30) Chloroform	5.23	83	3576	0.69	ug/l	95
38) Carbon Tetrachloride	5.78	117	10150	2.56	ug/l	91
44) Trichloroethene	7.21	130	23408	7.04	ug/l	98
64) Tetrachloroethene	9.34	164	9530	2.92	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081518\
Data File : VX004088.D
Acq On : 15 Aug 2018 22:14
Operator : JC/MD
Sample : J4439-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-295-SRI3

Quant Time: Aug 16 17:00:04 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 13:27:25 2018
Response via : Initial Calibration

