

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081518\
 Data File : VX004099.D
 Acq On : 16 Aug 2018 04:01
 Operator : JC/MD
 Sample : J4495-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS050MW109-180814

Quant Time: Aug 16 16:44:11 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 13:27:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	247582	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	389064	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	363436	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	195937	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	185848	53.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.56%	
35) Dibromofluoromethane	5.51	113	158969	54.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.18%	
50) Toluene-d8	8.72	98	557116	47.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.38%	
62) 4-Bromofluorobenzene	11.14	95	186230	43.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.98%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
4) Vinyl Chloride	1.40	62	24738	7.73	ug/l	# 78
21) trans-1,2-Dichloroethene	3.17	96	1624	0.55	ug/l	85
27) cis-1,2-Dichloroethene	4.60	96	28513	8.57	ug/l	90
52) Toluene	8.79	92	8134	0.97	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081518\
Data File : VX004099.D
Acq On : 16 Aug 2018 04:01
Operator : JC/MD
Sample : J4495-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS050MW109-180814

Quant Time: Aug 16 16:44:11 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 13:27:25 2018
Response via : Initial Calibration

