

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081518\
 Data File : VX004101.D
 Acq On : 16 Aug 2018 04:54
 Operator : JC/MD
 Sample : J4495-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS050MW612-180814

Quant Time: Aug 16 16:46:39 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 13:27:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	257550	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	404534	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	400795	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	227570	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	195637	53.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.82%	
35) Dibromofluoromethane	5.51	113	158911	52.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.98%	
50) Toluene-d8	8.72	98	587649	47.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.74%	
62) 4-Bromofluorobenzene	11.14	95	208209	46.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.54%	
Target Compounds						
4) Vinyl Chloride	1.41	62	14682	4.41	ug/l	# 81
27) cis-1,2-Dichloroethene	4.60	96	11743	3.39	ug/l	99
29) Tetrahydrofuran	5.17	42	6702	4.27	ug/l	95
52) Toluene	8.79	92	4185	0.48	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081518\
Data File : VX004101.D
Acq On : 16 Aug 2018 04:54
Operator : JC/MD
Sample : J4495-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS050MW612-180814

Quant Time: Aug 16 16:46:39 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 13:27:25 2018
Response via : Initial Calibration

