

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081518\
 Data File : VX004103.D
 Acq On : 16 Aug 2018 05:47
 Operator : JC/MD
 Sample : J4495-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS050MW636-180814

Quant Time: Aug 16 16:47:10 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 13:27:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	256495	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	402781	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	382705	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	213493	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	194099	53.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.42%	
35) Dibromofluoromethane	5.51	113	158510	52.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.16%	
50) Toluene-d8	8.72	98	577833	47.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.56%	
62) 4-Bromofluorobenzene	11.14	95	198470	44.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.54%	
Target Compounds						
4) Vinyl Chloride	1.40	62	88009	26.55	ug/l	98
27) cis-1,2-Dichloroethene	4.60	96	10184	2.96	ug/l	98
52) Toluene	8.79	92	4803	0.55	ug/l	95
65) Chlorobenzene	10.14	112	4552	0.46	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081518\
Data File : VX004103.D
Acq On : 16 Aug 2018 05:47
Operator : JC/MD
Sample : J4495-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS050MW636-180814

Quant Time: Aug 16 16:47:10 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 13:27:25 2018
Response via : Initial Calibration

