

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX081518\
 Data File : VX004105.D
 Acq On : 16 Aug 2018 06:40
 Operator : JC/MD
 Sample : J4495-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS050MW504-180814

Quant Time: Aug 16 16:47:42 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 13:27:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	257446	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	401198	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	376211	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	202070	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	193532	53.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.70%	
35) Dibromofluoromethane	5.51	113	156450	52.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.20%	
50) Toluene-d8	8.72	98	572535	47.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.06%	
62) 4-Bromofluorobenzene	11.14	95	192365	43.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.14%	
Target Compounds						
4) Vinyl Chloride	1.41	62	10813	3.25	ug/l	# 86
27) cis-1,2-Dichloroethene	4.60	96	14875	4.30	ug/l	95
44) Trichloroethene	7.22	130	1570	0.42	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX081518\
Data File : VX004105.D
Acq On : 16 Aug 2018 06:40
Operator : JC/MD
Sample : J4495-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS050MW504-180814

Quant Time: Aug 16 16:47:42 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 13:27:25 2018
Response via : Initial Calibration

