

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX081518\
 Data File : VX004112.D
 Acq On : 16 Aug 2018 09:46
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 16 14:22:50 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 13:27:25 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	93	0.00
2 T	Dichlorodifluoromethane	0.487	0.487	0.0	92	0.00
3 P	Chloromethane	0.597	0.571	4.4	93	0.00
4 C	Vinyl Chloride	0.646	0.630	2.5#	93	0.00
5 T	Bromomethane	0.369	0.343	7.0	99	0.00
6 T	Chloroethane	0.496	0.431	13.1	93	0.00
7 T	Trichlorofluoromethane	0.907	0.885	2.4	94	0.00
8 T	Diethyl Ether	0.396	0.393	0.8	97	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.566	0.532	6.0	91	0.00
10 T	Methyl Iodide	0.548	0.616	-12.4	101	0.00
11 T	Tert butyl alcohol	0.142	0.129	9.2	85	0.00
12 CM	1,1-Dichloroethene	0.526	0.518	1.5#	94	0.00
13 T	Acrolein	0.088	0.073	17.0	71	0.00
14 T	Allyl chloride	1.001	0.893	10.8	86	0.00
15 T	Acrylonitrile	0.354	0.340	4.0	92	0.00
16 T	Acetone	0.366	0.349	4.6	93	0.00
17 T	Carbon Disulfide	1.492	1.424	4.6	92	0.00
18 T	Methyl Acetate	0.814	0.798	2.0	97	0.00
19 T	Methyl tert-butyl Ether	1.853	1.833	1.1	96	0.00
20 T	Methylene Chloride	0.662	0.609	8.0	95	0.00
21 T	trans-1,2-Dichloroethene	0.597	0.573	4.0	93	0.00
22 T	Diisopropyl ether	1.961	2.065	-5.3	103	0.00
23 T	Vinyl Acetate	1.611	1.704	-5.8	98	0.00
24 P	1,1-Dichloroethane	1.111	1.145	-3.1	100	0.00
25 T	2-Butanone	0.596	0.571	4.2	92	0.00
26 T	2,2-Dichloropropane	0.883	0.365	58.7#	40#	0.00
27 T	cis-1,2-Dichloroethene	0.672	0.676	-0.6	98	0.00
28 T	Bromochloromethane	0.510	0.537	-5.3	97	0.00
29 T	Tetrahydrofuran	0.304	0.293	3.6	93	0.00
30 C	Chloroform	1.084	1.106	-2.0#	99	0.00
31 T	Cyclohexane	0.974	0.960	1.4	96	0.00
32 T	1,1,1-Trichloroethane	0.906	0.941	-3.9	99	0.00
33 S	1,2-Dichloroethane-d4	0.704	0.698	0.9	92	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	95	0.00
35 S	Dibromofluoromethane	0.374	0.389	-4.0	97	0.00
36 T	1,1-Dichloropropene	0.569	0.551	3.2	96	0.00
37 T	Ethyl Acetate	0.620	0.586	5.5	93	0.00
38 T	Carbon Tetrachloride	0.553	0.548	0.9	97	0.00
39 T	Methylcyclohexane	0.645	0.625	3.1	91	0.00
40 TM	Benzene	1.708	1.711	-0.2	98	0.00
41 T	Methacrylonitrile	0.347	0.324	6.6	94	0.00
42 TM	1,2-Dichloroethane	0.598	0.585	2.2	96	0.00
43 T	Isopropyl Acetate	0.913	0.904	1.0	95	0.00
44 TM	Trichloroethene	0.464	0.450	3.0	95	0.00
45 C	1,2-Dichloropropane	0.437	0.438	-0.2#	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX081518\
 Data File : VX004112.D
 Acq On : 16 Aug 2018 09:46
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Aug 16 14:22:50 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081418W.M
 Quant Title : SW846 8260
 QLast Update : Tue Aug 14 13:27:25 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.286	0.284	0.7	97	0.00
47 T	Bromodichloromethane	0.541	0.548	-1.3	96	0.00
48 T	Methyl methacrylate	0.460	0.478	-3.9	95	0.00
49 T	1,4-Dioxane	0.010	0.010	0.0	84	0.00
50 S	Toluene-d8	1.517	1.504	0.9	91	0.00
51 T	4-Methyl-2-Pentanone	0.749	0.766	-2.3	94	0.00
52 CM	Toluene	1.080	1.089	-0.8#	96	0.00
53 T	t-1,3-Dichloropropene	0.585	0.544	7.0	85	0.00
54 T	cis-1,3-Dichloropropene	0.648	0.603	6.9	86	0.00
55 T	1,1,2-Trichloroethane	0.446	0.442	0.9	97	0.00
56 T	Ethyl methacrylate	0.607	0.649	-6.9	95	0.00
57 T	1,3-Dichloropropane	0.740	0.744	-0.5	97	0.00
58 T	2-Chloroethyl Vinyl ether	0.309	0.338	-9.4	95	0.00
59 T	2-Hexanone	0.564	0.582	-3.2	94	0.00
60 T	Dibromochloromethane	0.429	0.442	-3.0	96	0.00
61 T	1,2-Dibromoethane	0.455	0.454	0.2	96	0.00
62 S	4-Bromofluorobenzene	0.550	0.517	6.0	86	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	87	0.00
64 T	Tetrachloroethene	0.474	0.536	-13.1	103	0.00
65 PM	Chlorobenzene	1.292	1.298	-0.5	91	0.00
66 T	1,1,1,2-Tetrachloroethane	0.441	0.467	-5.9	95	0.00
67 C	Ethyl Benzene	2.045	2.228	-8.9#	94	0.00
68 T	m/p-Xylenes	0.801	0.868	-8.4	92	0.00
69 T	o-Xylene	0.765	0.837	-9.4	90	0.00
70 T	Styrene	1.288	1.437	-11.6	92	0.00
71 P	Bromoform	0.349	0.365	-4.6	90	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	0.00
73 T	Isopropylbenzene	3.384	3.505	-3.6	93	0.00
74 T	N-amyl acetate	1.591	1.553	2.4	95	0.00
75 P	1,1,2,2-Tetrachloroethane	1.182	1.093	7.5	90	0.00
76 T	1,2,3-Trichloropropane	1.003	0.950	5.3	89	0.00
77 T	Bromobenzene	0.933	0.865	7.3	88	0.00
78 T	n-propylbenzene	3.787	3.824	-1.0	88	0.00
79 T	2-Chlorotoluene	2.308	2.207	4.4	88	0.00
80 T	1,3,5-Trimethylbenzene	2.771	2.817	-1.7	90	0.00
81 T	trans-1,4-Dichloro-2-butene	0.324	0.250	22.8#	69	0.00
82 T	4-Chlorotoluene	2.728	2.612	4.3	87	0.00
83 T	tert-Butylbenzene	2.793	2.874	-2.9	93	0.00
84 T	1,2,4-Trimethylbenzene	2.869	2.976	-3.7	91	0.00
85 T	sec-Butylbenzene	3.269	3.277	-0.2	89	0.00
86 T	p-Isopropyltoluene	2.960	3.126	-5.6	94	0.00
87 T	1,3-Dichlorobenzene	1.779	1.752	1.5	96	0.00
88 T	1,4-Dichlorobenzene	1.831	1.761	3.8	95	0.00
89 T	n-Butylbenzene	2.332	2.476	-6.2	99	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX081518\
Data File : VX004112.D
Acq On : 16 Aug 2018 09:46
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Aug 16 14:22:50 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X081418W.M
Quant Title : SW846 8260
QLast Update : Tue Aug 14 13:27:25 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.443	0.432	2.5	93	0.00
91 T	1,2-Dichlorobenzene	1.604	1.561	2.7	97	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.208	0.204	1.9	92	0.00
93 T	1,2,4-Trichlorobenzene	1.097	1.129	-2.9	92	0.00
94 T	Hexachlorobutadiene	0.550	0.500	9.1	89	0.00
95 T	Naphthalene	3.155	3.634	-15.2	96	0.00
96 T	1,2,3-Trichlorobenzene	1.196	1.243	-3.9	94	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6