

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081522\
 Data File : VX030623.D
 Acq On : 15 Aug 2022 13:14
 Operator : JC/MD
 Sample : N4110-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VHBLK001

Quant Time: Aug 16 04:29:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML080422WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Aug 16 04:27:08 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.769	114	259080	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	234171	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	103433	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.368	65	95104	57.162	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	114.320%
7) Chloroethane-d5	1.666	69	71230	60.731	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	121.460%
11) 1,1-Dichloroethene-d2	2.307	63	150589	47.849	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	95.700%
21) 2-Butanone-d5	4.471	46	166966	108.309	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	108.310%
24) Chloroform-d	5.068	84	189255	55.258	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	110.520%
26) 1,2-Dichloroethane-d4	5.964	65	135158	58.119	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	116.240%
32) Benzene-d6	5.977	84	372212	55.490	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	110.980%
36) 1,2-Dichloropropane-d6	7.312	67	119976	55.129	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	110.260%
41) Toluene-d8	8.653	98	334147	52.848	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	105.700%
43) trans-1,3-Dichloroprop...	8.952	79	54227	50.087	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	100.180%
47) 2-Hexanone-d5	9.391	63	110053	99.210	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	99.210%
56) 1,1,2,2-Tetrachloroeth...	11.195	84	151547	53.008	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	106.020%
66) 1,2-Dichlorobenzene-d4	12.323	152	109201	54.432	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	108.860%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX081522\
Data File : VX030623.D
Acq On : 15 Aug 2022 13:14
Operator : JC/MD
Sample : N4110-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VHBLK001

Quant Time: Aug 16 04:29:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML080422WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Aug 16 04:27:08 2022
Response via : Initial Calibration

